


REV.	DESCRIPTION	REVISIONS	DATE	APPROVED



EXP. CHAMBER HAS 2 X 50' DELAY PATHS. 50' PROVIDES
 2 TAU TIME DELAY FOR 3 CM/SEC FLOW IN THE EXP. CHAMBER
 FOOTPRINT OF RUBBER SKIRTS - ~140" X 82" OR ~30" WIDE EACH SIDE OF ENCLOSURE

Shop / Q.C.		UNLESS OTHERWISE SPECIFIED TOLERANCES ARE:		APPROVALS		DATE	Monterey Bay Aquarium Research Institute	
Inspector		1 PLACE DIM ±.174	2 SPACE DIM ±.005	3 PLACE DIM ±.005	ANGLE ±"		DRAWN	J SCHOLFIELD
Shop Foreman		1. DIMENSIONS ARE IN INCHES.					ENGINEER	
Engineer		2. DIMENSION LINES HELD AFTER PLATING.				4/23/14	CHECKED	
NEXT ASSY Description	NEXT ASSY Number	3. REMOVE BURRS AND SHARP EDGES .015 MAX.				PROJECT No.	901219.99	
APPLICATION		4. MACHINED FILLET RADI .015-.020.				PROJECT Name	shFOCE	
		5. MACHINED DIAMETERS ON A COMMON CENTERLINE CONCENTRIC WITHIN .005				WEIGHT		
		6. PERPENDICULARITY & PARALLELISM OF MACHINED SURFACES .002 PER INCH. MAX. .012 INCHES FOR SINGLE SURFACE.				SCALE	DWG. NO.	100XXXDWG
		DO NOT SCALE DRAWING				SCALE	CAD FILE:	SHEET 1 OF 1

